

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030824\
 Data File : VN081349.D
 Acq On : 08 Mar 2024 16:29
 Operator : JC\MD
 Sample : P1724-05 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 381

Quant Time: Mar 09 00:59:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 06 03:12:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	282938	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	529471	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	441133	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	162302	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	218269	53.354	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.700%	
35) Dibromofluoromethane	8.165	113	168340	51.894	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.780%	
50) Toluene-d8	10.570	98	614417	50.668	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.340%	
62) 4-Bromofluorobenzene	12.853	95	183095	42.720	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 85.440%	
Target Compounds						
16) Acetone	4.436	43	3867	2.668	ug/l	Qvalue # 69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030824\
Data File : VN081349.D
Acq On : 08 Mar 2024 16:29
Operator : JC\MD
Sample : P1724-05 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
381

Quant Time: Mar 09 00:59:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 06 03:12:57 2024
Response via : Initial Calibration

