

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN030920\
 Data File : VN060396.D
 Acq On : 9 Mar 2020 12:32
 Operator : JC/MD
 Sample : PB127354ZHE#01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127354ZHE#01

Quant Time: Mar 09 17:06:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	203868	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	336307	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	304996	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	132508	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	130529	49.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.80%	
35) Dibromofluoromethane	7.56	113	96259	61.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	123.32%	
50) Toluene-d8	10.08	98	409952	51.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.78%	
62) 4-Bromofluorobenzene	12.40	95	145997	48.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.50%	
Target Compounds						
20) Methylene Chloride	4.52	84	5658	2.44	ug/l #	91
43) Isopropyl Acetate	8.14	43	44305	8.34	ug/l #	89
80) 1,3,5-Trimethylbenzene	12.73	105	4309	0.52	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN030920\
Data File : VN060396.D
Acq On : 9 Mar 2020 12:32
Operator : JC/MD
Sample : PB127354ZHE#01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127354ZHE#01

Quant Time: Mar 09 17:06:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

