

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN030920\
 Data File : VN060397.D
 Acq On : 9 Mar 2020 12:55
 Operator : JC/MD
 Sample : PB127354ZHE#02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127354ZHE#02

Quant Time: Mar 09 17:08:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	205557	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	329910	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	307372	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	137478	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	131246	49.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.52%	
35) Dibromofluoromethane	7.56	113	96009	62.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	125.40%	
50) Toluene-d8	10.08	98	409195	52.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.60%	
62) 4-Bromofluorobenzene	12.40	95	148424	50.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.04%	
Target Compounds						
16) Acetone	3.79	43	5705	6.70	ug/l	93
20) Methylene Chloride	4.52	84	5289	2.26	ug/l	96
43) Isopropyl Acetate	8.14	43	43687	8.38	ug/l #	90
80) 1,3,5-Trimethylbenzene	12.73	105	4356	0.51	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN030920\
Data File : VN060397.D
Acq On : 9 Mar 2020 12:55
Operator : JC/MD
Sample : PB127354ZHE#02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127354ZHE#02

Quant Time: Mar 09 17:08:42 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

