

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN030920\
 Data File : VN060398.D
 Acq On : 9 Mar 2020 13:18
 Operator : JC/MD
 Sample : PB127354ZHE#03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127354ZHE#03

Quant Time: Mar 09 17:09:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	199667	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	326706	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	304706	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	136096	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	128268	50.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.14%	
35) Dibromofluoromethane	7.57	113	93877	61.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	123.82%	
50) Toluene-d8	10.08	98	403017	52.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.02%	
62) 4-Bromofluorobenzene	12.40	95	147433	50.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.34%	
Target Compounds						
16) Acetone	3.79	43	6030	7.29	ug/l	94
20) Methylene Chloride	4.53	84	5502	2.42	ug/l	99
43) Isopropyl Acetate	8.14	43	43173	8.36	ug/l #	91
80) 1,3,5-Trimethylbenzene	12.73	105	4382	0.52	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN030920\
Data File : VN060398.D
Acq On : 9 Mar 2020 13:18
Operator : JC/MD
Sample : PB127354ZHE#03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127354ZHE#03

Quant Time: Mar 09 17:09:28 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

