

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN030920\
 Data File : VN060399.D
 Acq On : 9 Mar 2020 13:41
 Operator : JC/MD
 Sample : PB127354ZHE#04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127354ZHE#04

Quant Time: Mar 09 17:12:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	200437	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	329032	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	308741	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	145521	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	131469	51.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.24%	
35) Dibromofluoromethane	7.57	113	94877	62.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	124.24%	
50) Toluene-d8	10.08	98	406556	52.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.20%	
62) 4-Bromofluorobenzene	12.40	95	152072	51.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.80%	
Target Compounds						
16) Acetone	3.79	43	5595	6.74	ug/l	95
20) Methylene Chloride	4.53	84	5381	2.36	ug/l	93
43) Isopropyl Acetate	8.14	43	43350	8.34	ug/l #	92
80) 1,3,5-Trimethylbenzene	12.73	105	4149	0.46	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN030920\
Data File : VN060399.D
Acq On : 9 Mar 2020 13:41
Operator : JC/MD
Sample : PB127354ZHE#04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127354ZHE#04

Quant Time: Mar 09 17:12:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

