

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030920\
 Data File : VN060404.D
 Acq On : 9 Mar 2020 15:56
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 09 18:21:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	91	0.00
2 T	Dichlorodifluoromethane	0.458	0.375	18.1	76	0.00
3 P	Chloromethane	0.798	0.671	15.9	81	0.00
4 C	Vinyl Chloride	0.684	0.543	20.6#	76	0.00
5 T	Bromomethane	0.301	0.205	31.9#	68	0.00
6 T	Chloroethane	0.371	0.314	15.4	82	0.00
7 T	Trichlorofluoromethane	0.875	0.755	13.7	82	0.00
8 T	Diethyl Ether	0.330	0.294	10.9	84	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.495	0.443	10.5	84	0.00
10 T	Methyl Iodide	0.452	0.360	20.4	75	0.00
11 T	Tert butyl alcohol	0.095	0.075	21.1	71	0.00
12 CM	1,1-Dichloroethene	0.525	0.428	18.5#	81	0.00
13 T	Acrolein	0.089	0.090	-1.1	88	0.00
14 T	Allyl chloride	0.886	0.745	15.9	77	0.00
15 T	Acrylonitrile	0.251	0.237	5.6	84	0.00
16 T	Acetone	0.207	0.187	9.7	84	0.00
17 T	Carbon Disulfide	1.490	1.065	28.5#	76	0.00
18 T	Methyl Acetate	0.691	0.607	12.2	86	0.00
19 T	Methyl tert-butyl Ether	1.744	1.635	6.2	88	0.00
20 T	Methylene Chloride	0.569	0.498	12.5	85	0.00
21 T	trans-1,2-Dichloroethene	0.595	0.469	21.2	83	0.00
22 T	Diisopropyl ether	1.794	1.749	2.5	91	0.00
23 T	Vinyl Acetate	1.426	1.377	3.4	87	0.00
24 P	1,1-Dichloroethane	1.016	0.956	5.9	88	0.00
25 T	2-Butanone	0.335	0.308	8.1	84	0.00
26 T	2,2-Dichloropropane	0.925	0.815	11.9	83	0.00
27 T	cis-1,2-Dichloroethene	0.660	0.579	12.3	88	0.00
28 T	Bromochloromethane	0.417	0.397	4.8	95	0.00
29 T	Tetrahydrofuran	0.233	0.209	10.3	83	0.00
30 C	Chloroform	1.011	0.966	4.5#	90	0.00
31 T	Cyclohexane	0.942	0.809	14.1	85	0.00
32 T	1,1,1-Trichloroethane	0.896	0.843	5.9	88	0.00
33 S	1,2-Dichloroethane-d4	0.642	0.617	3.9	92	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	91	0.00
35 S	Dibromofluoromethane	0.232	0.283	-22.0	114	0.00
36 T	1,1-Dichloropropene	0.477	0.424	11.1	87	0.00
37 T	Ethyl Acetate	0.451	0.418	7.3	85	0.00
38 T	Carbon Tetrachloride	0.464	0.423	8.8	85	0.00
39 T	Methylcyclohexane	0.562	0.488	13.2	84	0.00
40 TM	Benzene	1.387	1.293	6.8	88	0.00
41 T	Methacrylonitrile	0.208	0.222	-6.7	95	0.00
42 TM	1,2-Dichloroethane	0.500	0.478	4.4	89	0.00
43 T	Isopropyl Acetate	0.790	0.747	5.4	87	0.00
44 TM	Trichloroethene	0.386	0.348	9.8	87	0.00
45 C	1,2-Dichloropropane	0.363	0.360	0.8#	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030920\
 Data File : VN060404.D
 Acq On : 9 Mar 2020 15:56
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 09 18:21:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.235	0.221	6.0	87	0.00
47 T	Bromodichloromethane	0.479	0.478	0.2	92	0.00
48 T	Methyl methacrylate	0.360	0.330	8.3	85	0.00
49 T	1,4-Dioxane	0.005	0.004	20.0	78	0.00
50 S	Toluene-d8	1.175	1.118	4.9	92	0.00
51 T	4-Methyl-2-Pentanone	0.448	0.422	5.8	85	0.00
52 CM	Toluene	0.871	0.811	6.9#	87	0.00
53 T	t-1,3-Dichloropropene	0.586	0.546	6.8	89	0.00
54 T	cis-1,3-Dichloropropene	0.608	0.570	6.3	88	0.00
55 T	1,1,2-Trichloroethane	0.350	0.327	6.6	89	0.00
56 T	Ethyl methacrylate	0.503	0.497	1.2	87	0.00
57 T	1,3-Dichloropropane	0.571	0.554	3.0	89	0.00
58 T	2-Chloroethyl Vinyl ether	0.263	0.255	3.0	88	0.00
59 T	2-Hexanone	0.329	0.303	7.9	82	0.00
60 T	Dibromochloromethane	0.356	0.354	0.6	91	0.00
61 T	1,2-Dibromoethane	0.342	0.321	6.1	86	0.00
62 S	4-Bromofluorobenzene	0.445	0.437	1.8	94	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	93	0.00
64 T	Tetrachloroethene	0.444	0.372	16.2	85	0.00
65 PM	Chlorobenzene	1.010	0.908	10.1	87	0.00
66 T	1,1,1,2-Tetrachloroethane	0.373	0.351	5.9	89	0.00
67 C	Ethyl Benzene	1.839	1.704	7.3#	88	0.00
68 T	m/p-Xylenes	0.694	0.635	8.5	88	0.00
69 T	o-Xylene	0.661	0.618	6.5	89	0.00
70 T	Styrene	1.088	1.023	6.0	88	0.00
71 P	Bromoform	0.269	0.262	2.6	87	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
73 T	Isopropylbenzene	3.706	3.380	8.8	88	0.00
74 T	N-amyl acetate	1.619	1.459	9.9	87	0.00
75 P	1,1,2,2-Tetrachloroethane	1.007	0.957	5.0	91	0.00
76 T	1,2,3-Trichloropropane	1.097	0.873	20.4	77	0.00
77 T	Bromobenzene	0.908	0.815	10.2	88	0.00
78 T	n-propylbenzene	4.344	3.965	8.7	89	0.00
79 T	2-Chlorotoluene	2.554	2.306	9.7	89	0.00
80 T	1,3,5-Trimethylbenzene	3.120	2.862	8.3	88	0.00
81 T	trans-1,4-Dichloro-2-butene	0.397	0.353	11.1	82	0.00
82 T	4-Chlorotoluene	2.690	2.410	10.4	89	0.00
83 T	tert-Butylbenzene	2.701	2.400	11.1	86	0.00
84 T	1,2,4-Trimethylbenzene	3.149	2.873	8.8	88	0.00
85 T	sec-Butylbenzene	3.618	3.256	10.0	87	0.00
86 T	p-Isopropyltoluene	3.348	2.940	12.2	86	0.00
87 T	1,3-Dichlorobenzene	1.706	1.492	12.5	88	0.00
88 T	1,4-Dichlorobenzene	1.697	1.475	13.1	88	0.00
89 T	n-Butylbenzene	3.074	2.702	12.1	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030920\
Data File : VN060404.D
Acq On : 9 Mar 2020 15:56
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Mar 09 18:21:28 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.485	0.409	15.7	79	0.00
91 T	1,2-Dichlorobenzene	1.592	1.422	10.7	87	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.228	0.196	14.0	83	0.00
93 T	1,2,4-Trichlorobenzene	1.130	0.906	19.8	84	0.00
94 T	Hexachlorobutadiene	0.536	0.425	20.7	85	0.00
95 T	Naphthalene	3.258	2.532	22.3	79	0.00
96 T	1,2,3-Trichlorobenzene	1.062	0.834	21.5	82	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6