

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030921\
 Data File : VN066155.D
 Acq On : 9 Mar 2021 15:04
 Operator : JC/MD
 Sample : VN0309WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0309WBL01

Quant Time: Mar 10 02:55:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 09 14:48:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.584	168	196297	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	342448	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	298431	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	114987	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.949	65	185118	49.98	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.96%
35) Dibromofluoromethane	7.507	113	118445	50.70	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.40%
50) Toluene-d8	10.028	98	453162	49.30	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.60%
62) 4-Bromofluorobenzene	12.345	95	148345	43.35	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.70%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030921\
Data File : VN066155.D
Acq On : 9 Mar 2021 15:04
Operator : JC/MD
Sample : VN0309WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0309WBL01

Quant Time: Mar 10 02:55:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030921W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 09 14:48:21 2021
Response via : Initial Calibration

