

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030921\
 Data File : VN066160.D
 Acq On : 9 Mar 2021 17:32
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Mar 10 03:31:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 09 14:48:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.592	168	176597	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.518	114	303849	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	280590	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	116030	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.954	65	168244	50.49	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.98%	
35) Dibromofluoromethane	7.512	113	109016	52.59	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 105.18%	
50) Toluene-d8	10.030	98	411873	50.50	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 101.00%	
62) 4-Bromofluorobenzene	12.345	95	143804	47.36	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 94.72%	
Target Compounds						Qvalue
95) Naphthalene	15.072	128	12525	1.75	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030921\
Data File : VN066160.D
Acq On : 9 Mar 2021 17:32
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Mar 10 03:31:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030921W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 09 14:48:21 2021
Response via : Initial Calibration

