

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030922\
 Data File : VN071204.D
 Acq On : 09 Mar 2022 12:07
 Operator : JC\MD
 Sample : VN0309WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0309WBL01

Quant Time: Mar 10 00:56:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	849755	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1390087	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1250497	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	435924	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	532953	45.762	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.520%
35) Dibromofluoromethane	8.016	113	405757	46.256	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.520%
50) Toluene-d8	10.439	98	1687787	47.788	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.580%
62) 4-Bromofluorobenzene	12.727	95	582041	48.667	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030922\
Data File : VN071204.D
Acq On : 09 Mar 2022 12:07
Operator : JC\MD
Sample : VN0309WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0309WBL01

Quant Time: Mar 10 00:56:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

