

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030922\
 Data File : VN071210.D
 Acq On : 09 Mar 2022 15:21
 Operator : JC\MD
 Sample : N1830-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 350

Quant Time: Mar 10 01:12:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	710163	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1152279	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1051094	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	348586	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	451296	46.367	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.740%
35) Dibromofluoromethane	8.016	113	337677	46.440	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.880%
50) Toluene-d8	10.439	98	1428529	48.795	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.580%
62) 4-Bromofluorobenzene	12.727	95	492489	49.678	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.360%

Target Compounds						Qvalue
16) Acetone	4.293	43	70506	14.888	ug/l	97
18) Methyl Acetate	4.863	43	10554	1.292	ug/l	98
20) Methylene Chloride	5.104	84	30073	3.830	ug/l	95
25) 2-Butanone	7.340	43	14002	1.835	ug/l #	77
43) Isopropyl Acetate	8.557	43	85526	3.567	ug/l #	87
51) 4-Methyl-2-Pentanone	10.328	43	19603	1.317	ug/l #	85
93) 1,2,4-Trichlorobenzene	15.263	180	6493	4.058	ug/l	91
95) Naphthalene	15.504	128	22887	6.355	ug/l #	94
96) 1,2,3-Trichlorobenzene	15.698	180	6702	3.890	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030922\
Data File : VN071210.D
Acq On : 09 Mar 2022 15:21
Operator : JC\MD
Sample : N1830-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
350

Quant Time: Mar 10 01:12:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

