

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030922\
 Data File : VN071211.D
 Acq On : 09 Mar 2022 15:45
 Operator : JC\MD
 Sample : N1823-02DL 20X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DEWATERING-INFLUENT-030722DL

Quant Time: Mar 09 16:02:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	804269	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1316574	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1189962	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	406343	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	505535	55.025	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	110.060%
35) Dibromofluoromethane	8.016	113	379751	48.493	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.980%
50) Toluene-d8	10.439	98	1612555	51.276	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.560%
62) 4-Bromofluorobenzene	12.727	95	574409	51.099	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.200%
Target Compounds						
					Qvalue	
52) Toluene	10.498	92	1030076	43.195	ug/l	96
67) Ethyl Benzene	11.839	91	545410	12.153	ug/l	96
68) m/p-Xylenes	11.945	106	998034	57.572	ug/l	97
69) o-Xylene	12.274	106	260881	15.406	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030922\
Data File : VN071211.D
Acq On : 09 Mar 2022 15:45
Operator : JC\MD
Sample : N1823-02DL 20X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DEWATERING-INFLUENT-030722DL

Quant Time: Mar 09 16:02:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

