

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN030922\  
 Data File : VN071212.D  
 Acq On : 09 Mar 2022 16:09  
 Operator : JC\MD  
 Sample : N1835-05 10X  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 12 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 AUD-1272

Quant Time: Mar 10 01:13:02 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N030722W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Mar 08 06:52:46 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|--------|----------|
| -----                       |        |       |          |          |        |          |
| Internal Standards          |        |       |          |          |        |          |
| 1) Pentafluorobenzene       | 8.081  | 168   | 807305   | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 8.963  | 114   | 1327995  | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 11.739 | 117   | 1302581  | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.674 | 152   | 406705   | 50.000   | ug/l   | 0.00     |
| System Monitoring Compounds |        |       |          |          |        |          |
| 33) 1,2-Dichloroethane-d4   | 8.433  | 65    | 507642   | 45.881   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 61 - 141 | Recovery | =      | 91.760%  |
| 35) Dibromofluoromethane    | 8.016  | 113   | 379643   | 45.303   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 69 - 133 | Recovery | =      | 90.600%  |
| 50) Toluene-d8              | 10.439 | 98    | 1648196  | 48.849   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 65 - 126 | Recovery | =      | 97.700%  |
| 62) 4-Bromofluorobenzene    | 12.727 | 95    | 579182   | 50.692   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 58 - 135 | Recovery | =      | 101.380% |
| Target Compounds            |        |       |          |          |        |          |
|                             |        |       |          |          | Qvalue |          |
| 16) Acetone                 | 4.293  | 43    | 16657    | 3.094    | ug/l # | 89       |
| 18) Methyl Acetate          | 4.863  | 43    | 13348    | 1.385    | ug/l # | 78       |
| 20) Methylene Chloride      | 5.087  | 84    | 11322    | 1.524    | ug/l   | 90       |
| 52) Toluene                 | 10.504 | 92    | 7368     | 0.296    | ug/l   | 90       |
| 68) m/p-Xylenes             | 11.957 | 106   | 5935     | 0.328    | ug/l   | 97       |
| -----                       |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN030922\  
Data File : VN071212.D  
Acq On : 09 Mar 2022 16:09  
Operator : JC\MD  
Sample : N1835-05 10X  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 12 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
AUD-1272

Quant Time: Mar 10 01:13:02 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N030722W.M  
Quant Title : SW846 8260  
QLast Update : Tue Mar 08 06:52:46 2022  
Response via : Initial Calibration

