

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030922\
 Data File : VN071215.D
 Acq On : 09 Mar 2022 17:21
 Operator : JC\MD
 Sample : PB143182ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143182ZHE#01

Quant Time: Mar 10 00:58:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	858713	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1408809	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1296065	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	451226	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	542269	46.076	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.160%
35) Dibromofluoromethane	8.022	113	409261	46.036	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.080%
50) Toluene-d8	10.439	98	1749194	48.868	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.740%
62) 4-Bromofluorobenzene	12.727	95	605172	49.929	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.860%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	47784	8.344	ug/l #	87
18) Methyl Acetate	4.869	43	12856	1.298	ug/l #	77
20) Methylene Chloride	5.098	84	52829	5.391	ug/l #	86
43) Isopropyl Acetate	8.557	43	92624	3.160	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030922\
Data File : VN071215.D
Acq On : 09 Mar 2022 17:21
Operator : JC\MD
Sample : PB143182ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143182ZHE#01

Quant Time: Mar 10 00:58:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

