

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030923\
 Data File : VN076908.D
 Acq On : 09 Mar 2023 17:39
 Operator : JC\MD
 Sample : PB151321ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB151321ZHE#01

Quant Time: Mar 10 03:48:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 07 08:45:33 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	427482	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	734586	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	644510	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	267647	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	290806	48.521	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.040%
35) Dibromofluoromethane	8.172	113	222659	48.452	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.900%
50) Toluene-d8	10.572	98	875336	49.441	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.880%
62) 4-Bromofluorobenzene	12.854	95	311426	50.763	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.520%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	46423	17.745	ug/l	91
20) Methylene Chloride	5.295	84	18954	3.105	ug/l #	85
37) Ethyl Acetate	7.572	43	84206	11.107	ug/l	99
43) Isopropyl Acetate	8.695	43	371431	32.003	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030923\
Data File : VN076908.D
Acq On : 09 Mar 2023 17:39
Operator : JC\MD
Sample : PB151321ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB151321ZHE#01

Quant Time: Mar 10 03:48:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030623W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 07 08:45:33 2023
Response via : Initial Calibration

