

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030923\
 Data File : VN076910.D
 Acq On : 09 Mar 2023 18:28
 Operator : JC\MD
 Sample : PB151321ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB151321ZHE#03

Quant Time: Mar 10 03:49:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 07 08:45:33 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	421739	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	727396	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	640369	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	270045	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	290346	49.104	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.200%
35) Dibromofluoromethane	8.172	113	218152	47.941	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.880%
50) Toluene-d8	10.572	98	869365	49.589	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.180%
62) 4-Bromofluorobenzene	12.854	95	309297	50.914	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.820%
Target Compounds						
					Qvalue	
16) Acetone	4.443	43	44202	17.126	ug/l #	86
20) Methylene Chloride	5.296	84	19418	3.225	ug/l	93
37) Ethyl Acetate	7.572	43	81370	10.839	ug/l	98
43) Isopropyl Acetate	8.695	43	402809	35.050	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030923\
Data File : VN076910.D
Acq On : 09 Mar 2023 18:28
Operator : JC\MD
Sample : PB151321ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB151321ZHE#03

Quant Time: Mar 10 03:49:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030623W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 07 08:45:33 2023
Response via : Initial Calibration

