

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030923\
 Data File : VN076911.D
 Acq On : 09 Mar 2023 18:52
 Operator : JC\MD
 Sample : PB151321ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB151321ZHE#04

Quant Time: Mar 10 03:50:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 07 08:45:33 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	428795	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	744523	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	659543	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	274578	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	301282	50.115	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 100.220%		
35) Dibromofluoromethane	8.172	113	228090	48.972	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 97.940%		
50) Toluene-d8	10.571	98	893544	49.796	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 99.600%		
62) 4-Bromofluorobenzene	12.854	95	316104	50.838	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 101.680%		
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	44181	16.836	ug/l #	85
20) Methylene Chloride	5.295	84	19426	3.173	ug/l #	78
37) Ethyl Acetate	7.572	43	80859	10.523	ug/l	97
43) Isopropyl Acetate	8.695	43	432060	36.730	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030923\
Data File : VN076911.D
Acq On : 09 Mar 2023 18:52
Operator : JC\MD
Sample : PB151321ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB151321ZHE#04

Quant Time: Mar 10 03:50:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030623W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 07 08:45:33 2023
Response via : Initial Calibration

