

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031020\
 Data File : VN060410.D
 Acq On : 10 Mar 2020 12:45
 Operator : JC/MD
 Sample : L1857-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP

Quant Time: Mar 11 07:00:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	193168	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	313768	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	283603	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	124256	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	122076	49.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.52%	
35) Dibromofluoromethane	7.56	113	91570	62.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	125.74%	
50) Toluene-d8	10.08	98	380770	51.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.32%	
62) 4-Bromofluorobenzene	12.40	95	135461	48.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.96%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031020\
Data File : VN060410.D
Acq On : 10 Mar 2020 12:45
Operator : JC/MD
Sample : L1857-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP

Quant Time: Mar 11 07:00:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

