

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031020\
 Data File : VN060412.D
 Acq On : 10 Mar 2020 13:30
 Operator : JC/MD
 Sample : L1857-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-2

Quant Time: Mar 11 08:06:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	206072	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	336824	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	313751	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	144352	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	132652	50.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.34%	
35) Dibromofluoromethane	7.57	113	99430	63.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	127.20%	
50) Toluene-d8	10.08	98	414664	52.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.80%	
62) 4-Bromofluorobenzene	12.40	95	152926	50.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.96%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	3.74	101	17054	8.360	ug/l	95
27) cis-1,2-Dichloroethene	6.81	96	1266	0.466	ug/l	90
30) Chloroform	7.34	83	2432	0.584	ug/l	95
44) Trichloroethene	8.82	130	50956	19.582	ug/l	98
64) Tetrachloroethene	10.62	164	108641	43.803	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031020\
Data File : VN060412.D
Acq On : 10 Mar 2020 13:30
Operator : JC/MD
Sample : L1857-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-2

Quant Time: Mar 11 08:06:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

