

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031020\
 Data File : VN060415.D
 Acq On : 10 Mar 2020 14:38
 Operator : JC/MD
 Sample : L1857-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PW-1

Manual Integrations
 APPROVED

MMDadoda
 3/11/2020 3:23:13 PM

Quant Time: Mar 11 08:02:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	192083	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	311320	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	293356	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	134392	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	125357	50.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.74%	
35) Dibromofluoromethane	7.57	113	92433	63.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	127.94%	
50) Toluene-d8	10.08	98	386096	52.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.58%	
62) 4-Bromofluorobenzene	12.40	95	142805	51.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.02%	
Target Compounds						
4) Vinyl Chloride	2.16	62	38603	14.680	ug/l	98
19) Methyl tert-butyl Ether	5.01	73	9622m	1.436	ug/l	
27) cis-1,2-Dichloroethene	6.80	96	24731	9.758	ug/l	87
44) Trichloroethene	8.82	130	3688	1.533	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031020\
Data File : VN060415.D
Acq On : 10 Mar 2020 14:38
Operator : JC/MD
Sample : L1857-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PW-1

Manual Integrations
APPROVED

MMDadoda
3/11/2020 3:23:13 PM

Quant Time: Mar 11 08:02:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

