

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031020\
 Data File : VN060416.D
 Acq On : 10 Mar 2020 15:01
 Operator : JC/MD
 Sample : PB127392TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127392TB

Quant Time: Mar 11 07:06:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	190724	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	311964	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	292706	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	134017	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	122463	50.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.10%	
35) Dibromofluoromethane	7.57	113	91553	63.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	126.46%	
50) Toluene-d8	10.08	98	387464	52.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.74%	
62) 4-Bromofluorobenzene	12.40	95	143079	51.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.00%	
Target Compounds						
16) Acetone	3.79	43	6510	8.238	ug/l	93
20) Methylene Chloride	4.52	84	7759	3.575	ug/l	98
43) Isopropyl Acetate	8.14	43	57244	11.615	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031020\
Data File : VN060416.D
Acq On : 10 Mar 2020 15:01
Operator : JC/MD
Sample : PB127392TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127392TB

Quant Time: Mar 11 07:06:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

