

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031020\
 Data File : VN060421.D
 Acq On : 10 Mar 2020 16:54
 Operator : JC/MD
 Sample : PB127392ZHE#16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127392ZHE#16

Quant Time: Mar 11 07:28:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	204741	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	336869	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	321622	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	152118	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	133298	50.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.48%	
35) Dibromofluoromethane	7.57	113	99524	63.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	127.30%	
50) Toluene-d8	10.08	98	416050	52.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.14%	
62) 4-Bromofluorobenzene	12.40	95	157420	52.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.94%	
Target Compounds						
16) Acetone	3.79	43	6624	7.809	ug/l	92
20) Methylene Chloride	4.52	84	7773	3.337	ug/l	95
43) Isopropyl Acetate	8.14	43	57133	10.735	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031020\
Data File : VN060421.D
Acq On : 10 Mar 2020 16:54
Operator : JC/MD
Sample : PB127392ZHE#16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127392ZHE#16

Quant Time: Mar 11 07:28:13 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

