

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031020\  
 Data File : VN060426.D  
 Acq On : 10 Mar 2020 18:47  
 Operator : JC/MD  
 Sample : L1849-04  
 Misc : 5.00mL/MSVOA N/WATER  
 ALS Vial : 22 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VNJ-215-72-12018

Quant Time: Mar 11 07:40:48 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N022720W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Feb 27 13:52:50 2020  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units   | Dev(Min) |
|-----------------------------|-------|------|----------|--------|---------|----------|
| 1) Pentafluorobenzene       | 7.64  | 168  | 196235   | 50.00  | ug/l    | 0.00     |
| 34) 1,4-Difluorobenzene     | 8.57  | 114  | 319253   | 50.00  | ug/l    | 0.00     |
| 63) Chlorobenzene-d5        | 11.40 | 117  | 304039   | 50.00  | ug/l    | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.34 | 152  | 142333   | 50.00  | ug/l    | 0.00     |
| System Monitoring Compounds |       |      |          |        |         |          |
| 33) 1,2-Dichloroethane-d4   | 8.00  | 65   | 126459   | 50.23  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 100.46% |          |
| 35) Dibromofluoromethane    | 7.57  | 113  | 94547    | 63.80  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 127.60% |          |
| 50) Toluene-d8              | 10.08 | 98   | 396723   | 52.90  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 105.80% |          |
| 62) 4-Bromofluorobenzene    | 12.40 | 95   | 148549   | 52.25  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 104.50% |          |
| Target Compounds            |       |      |          |        |         |          |
| 16) Acetone                 | 3.79  | 43   | 13003    | 15.993 | ug/l    | 100      |
| 20) Methylene Chloride      | 4.52  | 84   | 7807     | 3.496  | ug/l    | 95       |
| 43) Isopropyl Acetate       | 8.14  | 43   | 53030    | 10.514 | ug/l #  | 91       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031020\  
Data File : VN060426.D  
Acq On : 10 Mar 2020 18:47  
Operator : JC/MD  
Sample : L1849-04  
Misc : 5.00mL/MSVOA N/WATER  
ALS Vial : 22 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VNJ-215-72-12018

Quant Time: Mar 11 07:40:48 2020  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_N\METHODS\82N022720W.M  
Quant Title : SW846 8260  
QLast Update : Thu Feb 27 13:52:50 2020  
Response via : Initial Calibration

