

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031020\
 Data File : VN060430.D
 Acq On : 10 Mar 2020 20:18
 Operator : JC/MD
 Sample : L1852-16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-3-S

Quant Time: Mar 11 07:48:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	200645	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	327332	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	311234	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	145699	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	128867	50.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.12%	
35) Dibromofluoromethane	7.57	113	96799	63.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	127.42%	
50) Toluene-d8	10.08	98	408677	53.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.30%	
62) 4-Bromofluorobenzene	12.40	95	152455	52.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.60%	
Target Compounds						
16) Acetone	3.79	43	7470	8.986	ug/l	98
20) Methylene Chloride	4.52	84	7710	3.377	ug/l	95
43) Isopropyl Acetate	8.14	43	49126	9.500	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031020\
Data File : VN060430.D
Acq On : 10 Mar 2020 20:18
Operator : JC/MD
Sample : L1852-16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3-S

Quant Time: Mar 11 07:48:47 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

