

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031020\
 Data File : VN060432.D
 Acq On : 10 Mar 2020 21:04
 Operator : JC/MD
 Sample : L1852-24
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-4-S

Quant Time: Mar 11 07:52:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	186737	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	306291	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	286118	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	130239	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	120943	50.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.96%	
35) Dibromofluoromethane	7.57	113	89752	63.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	126.26%	
50) Toluene-d8	10.08	98	377899	52.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.04%	
62) 4-Bromofluorobenzene	12.40	95	137453	50.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.78%	
Target Compounds						
16) Acetone	3.79	43	7462	9.645	ug/l	96
20) Methylene Chloride	4.53	84	8286	3.900	ug/l	96
43) Isopropyl Acetate	8.14	43	54854	11.336	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031020\
Data File : VN060432.D
Acq On : 10 Mar 2020 21:04
Operator : JC/MD
Sample : L1852-24
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-4-S

Quant Time: Mar 11 07:52:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

