

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031021\
 Data File : VN066166.D
 Acq On : 10 Mar 2021 15:14
 Operator : JC/MD
 Sample : MDL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MDL01

Manual Integrations
 APPROVED

MMDadoda
 3/11/2021 3:35:53 PM

Quant Time: Mar 10 17:54:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030921W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 10 17:54:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.582	168	171744	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.512	114	290218	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	249371	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	98631	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.947	65	177446	54.75	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 109.50%			
35) Dibromofluoromethane	7.509	113	109037	55.07	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 110.14%			
50) Toluene-d8	10.028	98	407487	52.31	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 104.62%			
62) 4-Bromofluorobenzene	12.345	95	127002	43.80	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 87.60%			
Target Compounds					Qvalue	
16) Acetone	3.744	43	4534	2.77	ug/l #	75
20) Methylene Chloride	4.457	84	1392m	0.58	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031021\
Data File : VN066166.D
Acq On : 10 Mar 2021 15:14
Operator : JC/MD
Sample : MDL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MDL01

Manual Integrations
APPROVED

MMDadoda
3/11/2021 3:35:53 PM

Quant Time: Mar 10 17:54:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030921W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 10 17:54:20 2021
Response via : Initial Calibration

