

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031021\
 Data File : VN066167.D
 Acq On : 10 Mar 2021 15:43
 Operator : JC/MD
 Sample : MDL02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MDL02

Manual Integrations
 APPROVED

MMDadoda
 3/11/2021 3:35:54 PM

Quant Time: Mar 10 18:03:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030921W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 10 17:54:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.587	168	160264	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	273995	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	241583	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	97406	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.949	65	168112	55.59	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 111.18%			
35) Dibromofluoromethane	7.509	113	102256	54.70	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 109.40%			
50) Toluene-d8	10.028	98	382195	51.97	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 103.94%			
62) 4-Bromofluorobenzene	12.345	95	124398	45.44	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 90.88%			
Target Compounds						Qvalue
16) Acetone	3.757	43	3442m	2.25	ug/l	
20) Methylene Chloride	4.457	84	1254m	0.56	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031021\
Data File : VN066167.D
Acq On : 10 Mar 2021 15:43
Operator : JC/MD
Sample : MDL02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MDL02

Manual Integrations
APPROVED

MMDadoda
3/11/2021 3:35:54 PM

Quant Time: Mar 10 18:03:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030921W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 10 17:54:20 2021
Response via : Initial Calibration

