

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031021\
 Data File : VN066171.D
 Acq On : 10 Mar 2021 17:27
 Operator : JC/MD
 Sample : M1603-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 32-NORTH

Quant Time: Mar 11 01:34:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 09 14:48:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.590	168	189614	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.518	114	320229	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	274161	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	108211	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.952	65	209623	58.59	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	117.18%
35) Dibromofluoromethane	7.515	113	59356	27.17	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	54.34%#
50) Toluene-d8	10.031	98	471778	54.89	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	109.78%
62) 4-Bromofluorobenzene	12.348	95	149011	46.57	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.14%
Target Compounds						
					Qvalue	
16) Acetone	3.752	43	129110	71.42	ug/l #	87
20) Methylene Chloride	4.463	84	9522	3.57	ug/l #	62
25) 2-Butanone	6.767	43	17262	6.16	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031021\
Data File : VN066171.D
Acq On : 10 Mar 2021 17:27
Operator : JC/MD
Sample : M1603-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
32-NORTH

Quant Time: Mar 11 01:34:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030921W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 09 14:48:21 2021
Response via : Initial Calibration

