

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031022\
 Data File : VN071228.D
 Acq On : 10 Mar 2022 14:11
 Operator : JC\MD
 Sample : N1834-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1281

Quant Time: Mar 10 14:31:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	625475	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1110493	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1031471	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	451676	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	436048	61.029	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 122.060%			
35) Dibromofluoromethane	8.016	113	327188	49.535	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 99.060%			
50) Toluene-d8	10.439	98	1353500	51.025	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 102.060%			
62) 4-Bromofluorobenzene	12.727	95	527889	55.675	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 111.360%			
Target Compounds						Qvalue
20) Methylene Chloride	5.087	84	63034	8.814	ug/l #	92
52) Toluene	10.498	92	61256	3.045	ug/l	98
59) 2-Hexanone	11.092	43	847152	110.670	ug/l #	51
74) N-amyl acetate	12.386	43	22568	1.942	ug/l #	88
92) 1,2-Dibromo-3-Chloropr...	14.604	75	2122	1.134	ug/l	82
93) 1,2,4-Trichlorobenzene	15.262	180	15639	3.804	ug/l	98
95) Naphthalene	15.504	128	92837	9.128	ug/l	99
96) 1,2,3-Trichlorobenzene	15.692	180	20723	5.083	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031022\
Data File : VN071228.D
Acq On : 10 Mar 2022 14:11
Operator : JC\MD
Sample : N1834-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1281

Quant Time: Mar 10 14:31:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

