

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031022\
 Data File : VN071229.D
 Acq On : 10 Mar 2022 14:35
 Operator : JC\MD
 Sample : N1835-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1272

Quant Time: Mar 10 14:52:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	823036	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1361979	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1262751	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	481890	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	520118	55.322	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	110.640%
35) Dibromofluoromethane	8.016	113	393040	48.517	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.040%
50) Toluene-d8	10.439	98	1689137	51.920	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.840%
62) 4-Bromofluorobenzene	12.727	95	605233	52.046	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.100%
Target Compounds						
						Qvalue
16) Acetone	4.287	43	142555	27.532	ug/l	99
18) Methyl Acetate	4.857	43	48038	3.679	ug/l	95
20) Methylene Chloride	5.092	84	64805	6.887	ug/l	94
25) 2-Butanone	7.339	43	32402	4.704	ug/l	96
95) Naphthalene	15.503	128	6102	5.245	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031022\
Data File : VN071229.D
Acq On : 10 Mar 2022 14:35
Operator : JC\MD
Sample : N1835-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1272

Quant Time: Mar 10 14:52:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

