

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031023\
 Data File : VN076926.D
 Acq On : 10 Mar 2023 15:51
 Operator : JC\MD
 Sample : VN0310MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0310MBL01

Quant Time: Mar 10 16:50:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 10 13:47:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	408495	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	666092	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	588940	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	249021	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	277964	45.599	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.200%
35) Dibromofluoromethane	8.172	113	208267	46.915	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.820%
50) Toluene-d8	10.572	98	789856	48.277	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.560%
62) 4-Bromofluorobenzene	12.854	95	280313	46.841	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031023\
Data File : VN076926.D
Acq On : 10 Mar 2023 15:51
Operator : JC\MD
Sample : VN0310MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0310MBL01

Quant Time: Mar 10 16:50:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 10 13:47:39 2023
Response via : Initial Calibration

