

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031023\
 Data File : VN076934.D
 Acq On : 10 Mar 2023 19:13
 Operator : JC\MD
 Sample : 01871-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Mar 13 01:23:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 10 13:47:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	374306	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	605604	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	536714	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	223427	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	258674	46.310	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.620%
35) Dibromofluoromethane	8.172	113	188025	46.585	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.180%
50) Toluene-d8	10.572	98	720526	48.438	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.880%
62) 4-Bromofluorobenzene	12.854	95	256575	47.157	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031023\
Data File : VN076934.D
Acq On : 10 Mar 2023 19:13
Operator : JC\MD
Sample : 01871-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Mar 13 01:23:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 10 13:47:39 2023
Response via : Initial Calibration

