

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031023\
 Data File : VN076939.D
 Acq On : 10 Mar 2023 21:14
 Operator : JC\MD
 Sample : 01870-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RAW-WATER

Quant Time: Mar 13 01:25:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 10 13:47:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	426439	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	702250	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	624466	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	261577	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	310559	48.802	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.600%
35) Dibromofluoromethane	8.172	113	223941	47.848	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.700%
50) Toluene-d8	10.572	98	838635	48.619	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.240%
62) 4-Bromofluorobenzene	12.854	95	296802	47.043	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031023\
Data File : VN076939.D
Acq On : 10 Mar 2023 21:14
Operator : JC\MD
Sample : 01870-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RAW-WATER

Quant Time: Mar 13 01:25:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 10 13:47:39 2023
Response via : Initial Calibration

