

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031023\
 Data File : VN076942.D
 Acq On : 10 Mar 2023 22:27
 Operator : JC\MD
 Sample : 01833-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 46

Quant Time: Mar 13 02:34:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 10 13:47:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	365088	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	614993	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	553296	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	235406	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	263744	48.410	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.820%
35) Dibromofluoromethane	8.178	113	190490	46.475	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.960%
50) Toluene-d8	10.572	98	732110	48.466	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.940%
62) 4-Bromofluorobenzene	12.854	95	266442	48.223	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.440%
Target Compounds						
					Qvalue	
16) Acetone	4.454	43	33413	17.699	ug/l	97
20) Methylene Chloride	5.296	84	16334	3.365	ug/l	97
37) Ethyl Acetate	7.566	43	56352	11.116	ug/l	98
43) Isopropyl Acetate	8.695	43	271445	30.333	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031023\
Data File : VN076942.D
Acq On : 10 Mar 2023 22:27
Operator : JC\MD
Sample : 01833-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
46

Quant Time: Mar 13 02:34:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 10 13:47:39 2023
Response via : Initial Calibration

