

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031023\
 Data File : VN076944.D
 Acq On : 10 Mar 2023 23:15
 Operator : JC\MD
 Sample : 01833-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 11

Quant Time: Mar 13 01:27:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 10 13:47:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	373269	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	618389	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	544765	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	234683	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	270830	48.621	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.240%
35) Dibromofluoromethane	8.172	113	192779	46.776	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.560%
50) Toluene-d8	10.572	98	743550	48.953	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.900%
62) 4-Bromofluorobenzene	12.854	95	269391	48.489	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.980%
Target Compounds						
					Qvalue	
16) Acetone	4.454	43	42223	21.875	ug/l #	88
20) Methylene Chloride	5.290	84	15853	3.194	ug/l	95
37) Ethyl Acetate	7.572	43	58889	11.553	ug/l	99
43) Isopropyl Acetate	8.695	43	309274	34.370	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031023\
Data File : VN076944.D
Acq On : 10 Mar 2023 23:15
Operator : JC\MD
Sample : 01833-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
11

Quant Time: Mar 13 01:27:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 10 13:47:39 2023
Response via : Initial Calibration

