

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031023\
 Data File : VN076923.D
 Acq On : 10 Mar 2023 14:11
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN031023

Quant Time: Mar 10 14:28:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 10 13:47:39 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	106	0.00
2 T	Dichlorodifluoromethane	0.901	0.929	-3.1	104	0.00
3 P	Chloromethane	0.487	0.436	10.5	102	0.00
4 C	Vinyl Chloride	0.452	0.437	3.3#	106	0.00
5 T	Bromomethane	0.333	0.317	4.8	105	0.00
6 T	Chloroethane	0.287	0.259	9.8	106	0.00
7 T	Trichlorofluoromethane	1.174	1.141	2.8	105	0.00
8 T	Diethyl Ether	0.378	0.374	1.1	104	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.600	0.578	3.7	106	0.00
10 T	Methyl Iodide	0.628	0.670	-6.7	106	0.00
11 T	Tert butyl alcohol	0.104	0.106	-1.9	105	0.00
12 CM	1,1-Dichloroethene	0.553	0.530	4.2#	103	0.00
13 T	Acrolein	0.093	0.089	4.3	110	0.00
14 T	Allyl chloride	0.793	0.763	3.8	106	0.00
15 T	Acrylonitrile	0.293	0.286	2.4	103	0.00
16 T	Acetone	0.259	0.291	-12.4	126	0.00
17 T	Carbon Disulfide	1.531	1.473	3.8	105	0.00
18 T	Methyl Acetate	0.954	0.903	5.3	105	0.00
19 T	Methyl tert-butyl Ether	1.965	1.978	-0.7	106	0.00
20 T	Methylene Chloride	0.665	0.595	10.5	104	0.00
21 T	trans-1,2-Dichloroethene	0.586	0.567	3.2	105	0.00
22 T	Diisopropyl ether	1.754	1.708	2.6	105	0.00
23 T	Vinyl Acetate	1.080	1.113	-3.1	104	0.00
24 P	1,1-Dichloroethane	1.120	1.075	4.0	105	0.00
25 T	2-Butanone	0.368	0.376	-2.2	110	0.00
26 T	2,2-Dichloropropane	0.992	1.006	-1.4	110	0.00
27 T	cis-1,2-Dichloroethene	0.676	0.654	3.3	105	0.00
28 T	Bromochloromethane	0.405	0.376	7.2	106	0.00
29 T	Tetrahydrofuran	0.228	0.228	0.0	105	0.00
30 C	Chloroform	1.204	1.158	3.8#	106	0.00
31 T	Cyclohexane	1.071	0.965	9.9	102	0.00
32 T	1,1,1-Trichloroethane	1.129	1.115	1.2	106	0.00
33 S	1,2-Dichloroethane-d4	0.746	0.716	4.0	107	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	104	0.00
35 S	Dibromofluoromethane	0.333	0.324	2.7	108	0.00
36 T	1,1-Dichloropropene	0.515	0.527	-2.3	106	0.00
37 T	Ethyl Acetate	0.412	0.441	-7.0	105	0.00
38 T	Carbon Tetrachloride	0.603	0.610	-1.2	106	0.00
39 T	Methylcyclohexane	0.626	0.649	-3.7	105	0.00
40 TM	Benzene	1.519	1.481	2.5	104	0.00
41 T	Methacrylonitrile	0.245	0.252	-2.9	105	0.00
42 TM	1,2-Dichloroethane	0.593	0.573	3.4	104	0.00
43 T	Isopropyl Acetate	0.728	0.746	-2.5	105	0.00
44 TM	Trichloroethene	0.393	0.388	1.3	107	0.00
45 C	1,2-Dichloropropane	0.374	0.363	2.9#	104	0.00
46 T	Dibromomethane	0.261	0.259	0.8	105	0.00
47 T	Bromodichloromethane	0.554	0.565	-2.0	107	0.00
48 T	Methyl methacrylate	0.356	0.363	-2.0	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031023\
 Data File : VN076923.D
 Acq On : 10 Mar 2023 14:11
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN031023

Quant Time: Mar 10 14:28:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 10 13:47:39 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.007	0.0	107	0.00
50 S	Toluene-d8	1.228	1.217	0.9	106	0.00
51 T	4-Methyl-2-Pentanone	0.437	0.443	-1.4	105	0.00
52 CM	Toluene	0.935	0.945	-1.1#	106	0.00
53 T	t-1,3-Dichloropropene	0.552	0.593	-7.4	106	0.00
54 T	cis-1,3-Dichloropropene	0.606	0.625	-3.1	105	0.00
55 T	1,1,2-Trichloroethane	0.364	0.357	1.9	105	0.00
56 T	Ethyl methacrylate	0.552	0.580	-5.1	104	0.00
57 T	1,3-Dichloropropane	0.641	0.627	2.2	105	0.00
58 T	2-Chloroethyl Vinyl ether	0.176	0.179	-1.7	103	0.00
59 T	2-Hexanone	0.319	0.334	-4.7	108	0.00
60 T	Dibromochloromethane	0.401	0.413	-3.0	107	0.00
61 T	1,2-Dibromoethane	0.365	0.369	-1.1	105	0.00
62 S	4-Bromofluorobenzene	0.449	0.460	-2.4	106	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	101	0.00
64 T	Tetrachloroethene	0.405	0.432	-6.7	105	0.00
65 PM	Chlorobenzene	1.118	1.114	0.4	105	0.00
66 T	1,1,1,2-Tetrachloroethane	0.437	0.438	-0.2	107	0.00
67 C	Ethyl Benzene	2.057	2.173	-5.6#	107	0.00
68 T	m/p-Xylenes	0.773	0.803	-3.9	106	0.00
69 T	o-Xylene	0.750	0.783	-4.4	105	0.00
70 T	Styrene	1.212	1.285	-6.0	105	0.00
71 P	Bromoform	0.316	0.338	-7.0	106	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	4.455	4.511	-1.3	107	0.00
74 T	N-amyl acetate	1.475	1.536	-4.1	106	0.00
75 P	1,1,2,2-Tetrachloroethane	1.311	1.203	8.2	107	0.00
76 T	1,2,3-Trichloropropane	1.212	0.984	18.8	90	0.00
77 T	Bromobenzene	0.996	0.981	1.5	108	0.00
78 T	n-propylbenzene	4.961	5.252	-5.9	107	0.00
79 T	2-Chlorotoluene	3.217	3.228	-0.3	106	0.00
80 T	1,3,5-Trimethylbenzene	3.795	3.970	-4.6	106	0.00
81 T	trans-1,4-Dichloro-2-butene	0.371	0.397	-7.0	107	0.00
82 T	4-Chlorotoluene	3.040	3.178	-4.5	108	0.00
83 T	tert-Butylbenzene	3.224	3.329	-3.3	107	0.00
84 T	1,2,4-Trimethylbenzene	3.797	3.917	-3.2	106	0.00
85 T	sec-Butylbenzene	4.420	4.709	-6.5	107	0.00
86 T	p-Isopropyltoluene	3.569	3.933	-10.2	108	0.00
87 T	1,3-Dichlorobenzene	1.815	1.833	-1.0	107	0.00
88 T	1,4-Dichlorobenzene	1.821	1.812	0.5	106	0.00
89 T	n-Butylbenzene	2.978	3.367	-13.1	110	0.00
90 T	Hexachloroethane	0.686	0.690	-0.6	107	0.00
91 T	1,2-Dichlorobenzene	1.824	1.804	1.1	106	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.289	0.295	-2.1	104	0.00
93 T	1,2,4-Trichlorobenzene	1.014	1.060	-4.5	106	0.00
94 T	Hexachlorobutadiene	0.540	0.588	-8.9	107	0.00
95 T	Naphthalene	3.018	3.348	-10.9	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031023\
Data File : VN076923.D
Acq On : 10 Mar 2023 14:11
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN031023

Quant Time: Mar 10 14:28:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 10 13:47:39 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.984	1.046	-6.3	105	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6