

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031025\
 Data File : VN085916.D
 Acq On : 10 Mar 2025 13:22
 Operator : JC\MD
 Sample : VN0310WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0310WBL01

Quant Time: Mar 11 01:37:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	185188	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	337149	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	297684	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	112686	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	131135	54.619	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.240%
35) Dibromofluoromethane	8.165	113	111691	50.627	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.260%
50) Toluene-d8	10.565	98	376828	46.947	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.900%
62) 4-Bromofluorobenzene	12.847	95	118920	44.967	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	89.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031025\
Data File : VN085916.D
Acq On : 10 Mar 2025 13:22
Operator : JC\MD
Sample : VN0310WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0310WBL01

Quant Time: Mar 11 01:37:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

