

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031025\
 Data File : VN085921.D
 Acq On : 10 Mar 2025 17:59
 Operator : JC\MD
 Sample : Q1514-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB03062025

Quant Time: Mar 11 01:40:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	193780	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	353088	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	308684	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	123437	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	134422	53.506	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.020%	
35) Dibromofluoromethane	8.165	113	116906	50.599	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.200%	
50) Toluene-d8	10.565	98	401211	47.729	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 95.460%	
62) 4-Bromofluorobenzene	12.847	95	130172	47.000	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 94.000%	
Target Compounds						Qvalue
16) Acetone	4.430	43	7801	9.813	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031025\
Data File : VN085921.D
Acq On : 10 Mar 2025 17:59
Operator : JC\MD
Sample : Q1514-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB03062025

Quant Time: Mar 11 01:40:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

