

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031025\
 Data File : VN085923.D
 Acq On : 10 Mar 2025 18:47
 Operator : JC\MD
 Sample : Q1516-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GAS-AUD-25-0026

Quant Time: Mar 11 01:41:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	181631	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	336202	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	294281	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	120577	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	134756	57.227	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.460%
35) Dibromofluoromethane	8.165	113	113352	51.525	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.040%
50) Toluene-d8	10.565	98	388785	48.573	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.140%
62) 4-Bromofluorobenzene	12.847	95	125421	47.559	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	95.120%
Target Compounds						
16) Acetone	4.436	43	3333	4.473	ug/l	Qvalue # 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031025\
Data File : VN085923.D
Acq On : 10 Mar 2025 18:47
Operator : JC\MD
Sample : Q1516-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GAS-AUD-25-0026

Quant Time: Mar 11 01:41:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

