

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031120\
 Data File : VN060437.D
 Acq On : 11 Mar 2020 9:56
 Operator : JC/MD
 Sample : VN0311WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0311WBL01

Quant Time: Mar 12 03:51:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	205382	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	334174	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	306822	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	135868	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	130939	49.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.38%	
35) Dibromofluoromethane	7.56	113	98981	63.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	127.62%	
50) Toluene-d8	10.08	98	408504	52.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.08%	
62) 4-Bromofluorobenzene	12.40	95	147919	49.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.42%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031120\
Data File : VN060437.D
Acq On : 11 Mar 2020 9:56
Operator : JC/MD
Sample : VN0311WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0311WBL01

Quant Time: Mar 12 03:51:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

