

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031120\
 Data File : VN060447.D
 Acq On : 11 Mar 2020 14:04
 Operator : JC/MD
 Sample : L1859-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-7-SA

Quant Time: Mar 12 04:16:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	192509	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	318319	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	301439	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	141825	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	123513	50.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.02%	
35) Dibromofluoromethane	7.56	113	94757	64.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	128.26%	
50) Toluene-d8	10.08	98	395087	52.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.66%	
62) 4-Bromofluorobenzene	12.40	95	149094	52.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.20%	
Target Compounds						
16) Acetone	3.79	43	5002	6.271	ug/l	98
20) Methylene Chloride	4.53	84	17846	8.147	ug/l	99
43) Isopropyl Acetate	8.14	43	54738	10.885	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031120\
Data File : VN060447.D
Acq On : 11 Mar 2020 14:04
Operator : JC/MD
Sample : L1859-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-7-SA

Quant Time: Mar 12 04:16:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

