

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031120\
 Data File : VN060448.D
 Acq On : 11 Mar 2020 14:26
 Operator : JC/MD
 Sample : L1859-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-7-DA

Quant Time: Mar 12 04:18:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	201715	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	331258	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	308796	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	143699	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	129766	50.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.28%	
35) Dibromofluoromethane	7.57	113	97604	63.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	126.96%	
50) Toluene-d8	10.08	98	406778	52.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.54%	
62) 4-Bromofluorobenzene	12.40	95	151393	51.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.64%	
Target Compounds						
16) Acetone	3.79	43	5351	6.403	ug/l	92
20) Methylene Chloride	4.53	84	17524	7.635	ug/l	93
43) Isopropyl Acetate	8.14	43	55452	10.596	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031120\
Data File : VN060448.D
Acq On : 11 Mar 2020 14:26
Operator : JC/MD
Sample : L1859-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-7-DA

Quant Time: Mar 12 04:18:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

