

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031120\
 Data File : VN060456.D
 Acq On : 11 Mar 2020 17:28
 Operator : JC/MD
 Sample : L1873-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1

Quant Time: Mar 12 04:49:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	192973	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	320493	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	305070	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	141696	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	126507	51.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.20%	
35) Dibromofluoromethane	7.56	113	95410	64.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	128.28%	
50) Toluene-d8	10.08	98	392955	52.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.38%	
62) 4-Bromofluorobenzene	12.40	95	149574	52.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.82%	
Target Compounds						
16) Acetone	3.79	43	6294	7.872	ug/l	95
20) Methylene Chloride	4.53	84	14101	6.422	ug/l	91
43) Isopropyl Acetate	8.14	43	54180	10.701	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031120\
Data File : VN060456.D
Acq On : 11 Mar 2020 17:28
Operator : JC/MD
Sample : L1873-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1

Quant Time: Mar 12 04:49:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

