


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN031120\
 Data File : VN060434.D
 Acq On : 11 Mar 2020 8:24
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Title : SW846 8260
 Last Update : Thu Feb 27 13:52:50 2020



AutoFind: Scans 3303, 3304, 3305; Background Corrected with Scan 3290

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.2	6563	PASS
75	95	30	60	52.0	16872	PASS
95	95	100	100	100.0	32421	PASS
96	95	5	9	7.6	2457	PASS
173	174	0.00	2	0.8	203	PASS
174	95	50	100	80.7	26170	PASS
175	174	5	9	7.7	2011	PASS
176	174	95	101	96.0	25114	PASS
177	176	5	9	6.3	1588	PASS