

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031122\
 Data File : VN071241.D
 Acq On : 11 Mar 2022 10:29
 Operator : JC\MD
 Sample : VN0311MBL01
 Misc : 5.00g/10.0mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0311MBL01

Quant Time: Mar 12 02:15:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	706315	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1176186	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1057095	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	353610	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	460769	47.599	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.200%
35) Dibromofluoromethane	8.016	113	349978	47.153	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.300%
50) Toluene-d8	10.439	98	1416852	47.412	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.820%
62) 4-Bromofluorobenzene	12.727	95	437322	43.216	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031122\
Data File : VN071241.D
Acq On : 11 Mar 2022 10:29
Operator : JC\MD
Sample : VN0311MBL01
Misc : 5.00g/10.0mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0311MBL01

Quant Time: Mar 12 02:15:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

