

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031122\
 Data File : VN071250.D
 Acq On : 11 Mar 2022 16:34
 Operator : JC\MD
 Sample : N1886-01
 Misc : 1.05g/10.0mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 2822-D

Quant Time: Mar 12 02:17:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.075	168	750792	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.957	114	1226069	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1103748	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	490971	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.428	65	465965	45.284	ug/l	-0.01
Spiked Amount	50.000	Range	61 - 141	Recovery	=	90.560%
35) Dibromofluoromethane	8.010	113	343858	44.444	ug/l	-0.01
Spiked Amount	50.000	Range	69 - 133	Recovery	=	88.880%
50) Toluene-d8	10.433	98	1510135	48.478	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.960%
62) 4-Bromofluorobenzene	12.727	95	522111	49.496	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.000%
Target Compounds						
					Qvalue	
18) Methyl Acetate	4.851	43	67021	5.408	ug/l	91
25) 2-Butanone	7.340	43	80694	10.005	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031122\
Data File : VN071250.D
Acq On : 11 Mar 2022 16:34
Operator : JC\MD
Sample : N1886-01
Misc : 1.05g/10.0mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
2822-D

Quant Time: Mar 12 02:17:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

