

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031125\
 Data File : VN085929.D
 Acq On : 11 Mar 2025 13:56
 Operator : JC\MD
 Sample : VN0311MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0311MBL01

Quant Time: Mar 12 01:19:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	148406	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	269925	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	239958	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	95589	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	119185	61.945	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	123.900%
35) Dibromofluoromethane	8.165	113	101108	57.244	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	114.480%
50) Toluene-d8	10.565	98	334018	51.978	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.960%
62) 4-Bromofluorobenzene	12.847	95	98736	46.633	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	93.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031125\
Data File : VN085929.D
Acq On : 11 Mar 2025 13:56
Operator : JC\MD
Sample : VN0311MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0311MBL01

Quant Time: Mar 12 01:19:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

