

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031125\
 Data File : VN085933.D
 Acq On : 11 Mar 2025 15:52
 Operator : JC\MD
 Sample : PB167050TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167050TB

Quant Time: Mar 12 01:22:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	159816	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	301730	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	300368	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	113239	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	116926	56.433	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 112.860%	
35) Dibromofluoromethane	8.165	113	100242	50.771	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.540%	
50) Toluene-d8	10.565	98	369443	51.430	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.860%	
62) 4-Bromofluorobenzene	12.847	95	124450	52.582	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 105.160%	
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	30349	46.292	ug/l	99
20) Methylene Chloride	5.271	84	5938	2.816	ug/l #	76
43) Isopropyl Acetate	8.688	43	59676	14.398	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031125\
Data File : VN085933.D
Acq On : 11 Mar 2025 15:52
Operator : JC\MD
Sample : PB167050TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167050TB

Quant Time: Mar 12 01:22:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

