

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031125\
 Data File : VN085934.D
 Acq On : 11 Mar 2025 16:16
 Operator : JC\MD
 Sample : Q1514-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB03062025

Quant Time: Mar 12 01:23:43 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	176702	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	336260	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	299219	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	119405	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	134504	58.713	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	117.420%
35) Dibromofluoromethane	8.165	113	112664	51.203	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.400%
50) Toluene-d8	10.565	98	387134	48.359	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.720%
62) 4-Bromofluorobenzene	12.847	95	125306	47.507	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	95.020%
Target Compounds						
16) Acetone	4.442	43	7216	9.955	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031125\
Data File : VN085934.D
Acq On : 11 Mar 2025 16:16
Operator : JC\MD
Sample : Q1514-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB03062025

Quant Time: Mar 12 01:23:43 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

