

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031125\
 Data File : VN085938.D
 Acq On : 11 Mar 2025 17:52
 Operator : JC\MD
 Sample : PB167050ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167050ZHE#05

Quant Time: Mar 12 01:25:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	160603	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	306695	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	293575	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	104809	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	129530	62.209	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 124.420%	
35) Dibromofluoromethane	8.165	113	110168	54.896	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 109.800%	
50) Toluene-d8	10.565	98	401954	55.050	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 110.100%	
62) 4-Bromofluorobenzene	12.847	95	115833	48.149	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 96.300%	
Target Compounds						
						Qvalue
16) Acetone	4.430	43	29765	45.178	ug/l	97
20) Methylene Chloride	5.283	84	6351	2.998	ug/l	96
43) Isopropyl Acetate	8.689	43	67944	16.194	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031125\
Data File : VN085938.D
Acq On : 11 Mar 2025 17:52
Operator : JC\MD
Sample : PB167050ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167050ZHE#05

Quant Time: Mar 12 01:25:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

