

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031125\
 Data File : VN085940.D
 Acq On : 11 Mar 2025 18:40
 Operator : JC\MD
 Sample : PB167050ZHE#07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167050ZHE#07

Quant Time: Mar 12 01:27:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	151984	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	298839	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	278612	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	94621	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	125448	63.666	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 127.340%#		
35) Dibromofluoromethane	8.159	113	103759	53.061	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 106.120%		
50) Toluene-d8	10.565	98	385323	54.160	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 108.320%		
62) 4-Bromofluorobenzene	12.847	95	108665	46.357	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 92.720%		
Target Compounds						
						Qvalue
16) Acetone	4.430	43	27207	43.638	ug/l #	85
20) Methylene Chloride	5.271	84	5306	2.646	ug/l	92
43) Isopropyl Acetate	8.688	43	75220	18.475	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031125\
Data File : VN085940.D
Acq On : 11 Mar 2025 18:40
Operator : JC\MD
Sample : PB167050ZHE#07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167050ZHE#07

Quant Time: Mar 12 01:27:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

